



Ref. Certif. No

NO118386/A2/M3

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

## Product

Intraoral Scanner

## Name and address of the applicant

Medit Corp.  
9F, 10F, 13F, 14F, 16F, 8, Yangpyeong-ro 25-gil, Yeongdeungpo-gu, Seoul, 07207,  
Korea, Republic of

## Name and address of the manufacturer

Medit Corp.  
9F, 10F, 13F, 14F, 16F, 8, Yangpyeong-ro 25-gil, Yeongdeungpo-gu, Seoul, 07207,  
Korea, Republic of

## Name and address of the factory

Medit Corp.  
9F, 10F, 13F, 14F, 16F, 8, Yangpyeong-ro 25-gil, Yeongdeungpo-gu, Seoul, 07207,  
Korea, Republic of

Note: When more than one factory, please report on page 2

 Additional Information on Page 2

## Ratings and Principal characteristics

- Handpiece: 4 A, 4 Vdc or internally powered
- Battery charger: 5 A, 12 Vdc
- Wireless Hub: 5 A, 12 Vdc or 3 A, 5 Vdc / 9 Vdc
- Auto calibration tool: Internally powered
- AC adapter: (Input) 1.6-0.7 A, 100-240 V~, 50-60 Hz, Cl. I / (Output) 5 A, 12 Vdc

Medit Corp.

## Trademark /Brand (If any)

## Customer's Testing Facility (CTF) Stage used

## Model/Type Ref

MD-IS0300

## Additional information (If necessary may also be reported on Page 2)

Classification: See summary of testing in test report for details. In addition, all applicable clauses of IEC 62366:2007 + AMD1:2014 have been addressed.

 Additional Information on Page 2

## A sample of the product was tested and found to be in conformity with

IEC 60601-1-6:2010, IEC 60601-1-6:2010/AMD1:2013  
IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012  
IEC 80601-2-60:2019

## As shown in the Test Report Ref. No. which forms part of this Certificate

REP002922

This certificate replaces the certificate NO118386/A1/M2, due to update the address of applicant, manufacturer, factory and technical modification.

This CB Test Certificate is issued by the National Certification Body

Philip Pedersen vei 11,  
NO-1366 Lysaker, Norway  
Date: 12-01-2023Signature: Okhyun Jeon  
Certification department



Ref. Certif. No

NO125339

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

## Product

Intraoral Scanner

## Name and address of the applicant

Medit Corp.  
9F, 10F, 13F, 14F, 16F, 8, Yangpyeong-ro 25-gil, Yeongdeungpo-gu, Seoul, 07207,  
Korea, Republic of

## Name and address of the manufacturer

Medit Corp.  
9F, 10F, 13F, 14F, 16F, 8, Yangpyeong-ro 25-gil, Yeongdeungpo-gu, Seoul, 07207,  
Korea, Republic of

## Name and address of the factory

Medit Corp.  
9F, 10F, 13F, 14F, 16F, 8, Yangpyeong-ro 25-gil, Yeongdeungpo-gu, Seoul, 07207,  
Korea, Republic of

Note: When more than one factory, please report on page 2

 Additional Information on Page 2

## Ratings and Principal characteristics

- Handpiece: 4 A, 4 Vdc or internally powered
- Battery charger: 5 A, 12 Vdc
- Wireless Hub: 5 A, 12 Vdc or 3 A, 5 Vdc / 9 Vdc
- Auto calibration tool: Internally powered
- AC adapter: (Input) 1.6-0.7 A, 100-240 V~, 50-60 Hz, Cl. I / (Output) 5 A, 12 Vdc

Medit Corp.

## Trademark /Brand (If any)

## Customer's Testing Facility (CTF) Stage used

## Model/Type Ref

MD-IS0400

## Additional information (If necessary may also be reported on Page 2)

Classification: See summary of testing in test report for details. In addition, all applicable clauses of IEC 62366:2007 + AMD1:2014 have been addressed.

 Additional Information on Page 2

## A sample of the product was tested and found to be in conformity with

IEC 60601-1-6:2010, IEC 60601-1-6:2010/AMD1:2013  
IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012  
IEC 80601-2-60:2019

## As shown in the Test Report Ref. No. which forms part of this Certificate

REP002922

This CB Test Certificate is issued by the National Certification Body

Philip Pedersen vei 11,  
NO-1366 Lysaker, Norway  
Date: 12-01-2023Signature: Okhyun Jeon  
Certification department